

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.282 °C	15.7 °C	0.418 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.179 °C	24.8 °C	-0.379 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.218 °C	50.1 °C	-0.118 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.130 %HR	27.7 %HR	2.570 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.325 %HR	47.6 %HR	-2.725 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.253 %HR	77.2 %HR	1.947 %HR	± 5.0 %HR	± 1.9E+00 %HR