

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.982 °C	29.9 °C	-0.082 °C	± 0.3 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	28.6 %HR	-1.459 %HR	± 4.0 %HR	± 1.9E+00 %HR
50 %HR	50.233 %HR	48.9 %HR	-1.333 %HR	± 4.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68.4 %HR	-1.843 %HR	± 4.0 %HR	± 1.9E+00 %HR