

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
18 °C	18.029 °C	17.7 °C	-0.329 °C	± 1.6 °C	± 4.0E-01 °C
23 °C	23.244 °C	23.2 °C	-0.044 °C	± 1.6 °C	± 4.0E-01 °C
28 °C	28.057 °C	27.9 °C	-0.157 °C	± 1.6 °C	± 4.0E-01 °C
45 %HR	45.396 %HR	47.8 %HR	2.404 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	60.284 %HR	57.4 %HR	-2.884 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.356 %HR	75.1 %HR	-0.256 %HR	± 5.0 %HR	± 1.9E+00 %HR