

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
18 °C	18.029 °C	18.0 °C	-0.029 °C	± 1.6 °C	± 4.0E-01 °C
23 °C	23.244 °C	23.4 °C	0.156 °C	± 1.6 °C	± 4.0E-01 °C
28 °C	28.057 °C	28.1 °C	0.043 °C	± 1.6 °C	± 4.0E-01 °C
45 %HR	45.396 %HR	48.1 %HR	2.704 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	60.284 %HR	60.5 %HR	0.216 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.356 %HR	74.5 %HR	-0.86 %HR	± 5.0 %HR	± 1.9E+00 %HR