

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.75	0.750000	0.270	± 0.3	± 1.5E+00 mm					
0 a 0.8	0.800030	0.240	± 0.3	± 1.5E+00 mm					
0 a 0.9	0.900030	0.250	± 0.3	± 1.5E+00 mm					
0 a 1	1.000040	0.240	± 1.0	± 1.5E+00 mm					
0 a 1.2	1.200040	0.250	± 1.0	± 1.5E+00 mm					
0 a 1.4	1.400050	0.250	± 0.4	± 1.5E+00 mm					
0 a 1.6	1.600030	0.280	± 0.5	± 1.5E+00 mm					
0 a 1.85	1.850002	0.320	± 0.5	± 1.5E+00 mm					
0 a 2	2.000020	0.310	± 0.6	± 1.5E+00 mm					