

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
28 °C	27.937 °C	28.4 °C	0.463 °C	± 1.0 °C	± 4.0E-01 °C
25 %HR	24.629 %HR	20 %HR	-4.629 %HR	± 5.0 %HR	± 1.9E+00 %HR
45 %HR	45.683 %HR	41 %HR	-4.683 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.957 %HR	70 %HR	-4.957 %HR	± 5.0 %HR	± 1.9E+00 %HR