

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
180.243 s	180 s	-0.243 s	-0.1348%	± 0.0500%	±	5.0E-02 s
600.131 s	600 s	-0.131 s	-0.0218%	± 0.0500%	±	5.0E-02 s
1800.246 s	1800 s	-0.246 s	-0.0137%	± 0.0500%	±	5.2E-02 s
3600.389 s	3600 s	-0.389 s	-0.0108%	± 0.0500%	±	5.6E-02 s