

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1800.343 s	1800.34 s	-0.003 s	-0.0002%	± 0.00058%	± 9.2E-02 s
3600.252 s	3600.25 s	-0.002 s	-0.0001%	± 0.00058%	± 1.1E-01 s
21600.354 s	21600.35 s	-0.004 s	-0.00002%	± 0.00058%	± 2.6E-01 s
43200.354 s	43200.35 s	-0.004 s	-0.00001%	± 0.00058%	± 3.8E-01 s
64800.246 s	64800.24 s	-0.006 s	-0.00001%	± 0.00058%	± 5.1E-01 s
86400.332 s	86400.33 s	-0.002 s	-0.000002%	± 0.00058%	± 6.1E-01 s