

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.719 °C	10.6 °C	-0.119 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.045 °C	19.8 °C	-0.245 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.153 °C	29.8 °C	-0.353 °C	± 0.5 °C	± 4.0E-01 °C
10 %HR	11.736 %HR	9.6 %HR	-2.136 %HR	± 3.0 %HR	± 1.9E+00 %HR
30 %HR	31.787 %HR	29.8 %HR	-1.987 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	61.327 %HR	63.8 %HR	2.473 %HR	± 3.0 %HR	± 1.9E+00 %HR