

Certificado No. (Certificate #):

62883

Equipo (Instrument):

INDICADOR DE DIAL

Fecha de Calibración (Calibration Date):

2025-10-29

Marca (Brand):

Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.00 mm	1.00 mm	0 µm	± µm	±	5.8E+00 µm
3.00 mm	3.00 mm	0 µm	± µm	±	5.8E+00 µm
5.00 mm	5.00 mm	0 µm	± µm	±	5.8E+00 µm
8.00 mm	8.00 mm	0 µm	± µm	±	5.8E+00 µm
10.00 mm	10.00 mm	0 µm	± µm	±	5.8E+00 µm
12.00 mm	12.00 mm	0 µm	± µm	±	5.8E+00 µm
15.00 mm	15.00 mm	0 µm	± µm	±	5.8E+00 µm
17.00 mm	17.00 mm	0 µm	± µm	±	5.8E+00 µm
20.00 mm	20.00 mm	0 µm	± µm	±	5.8E+00 µm