

Certificado No. (Certificate #):

62884

Equipo (Instrument):

INDICADOR DE DIAL

Fecha de Calibración (Calibration Date):

2025-10-29

Marca (Brand):

Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.00 mm	1.02 mm	20 µm		±	5.8E+00 µm
3.00 mm	3.03 mm	30 µm		±	5.8E+00 µm
5.00 mm	5.04 mm	40 µm		±	5.8E+00 µm
7.00 mm	7.04 mm	40 µm		±	5.8E+00 µm
10.00 mm	10.04 mm	40 µm		±	5.8E+00 µm
13.00 mm	13.04 mm	40 µm		±	5.8E+00 µm
15.00 mm	15.04 mm	40 µm		±	5.8E+00 µm
18.00 mm	18.04 mm	40 µm		±	5.8E+00 µm
20.00 mm	20.04 mm	40 µm		±	5.8E+00 µm