

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.971 °C	15.2 °C	0.229 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.998 °C	25.3 °C	0.302 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.018 °C	32.4 °C	0.382 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	22.3 %HR	1.955 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.304 %HR	52.4 %HR	2.096 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.131 %HR	71.3 %HR	1.169 %HR	± 5.0 %HR	± 1.9E+00 %HR