

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 15 °C                      | 15.055 °C                  | 15.3 °C            | 0.245 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 24.991 °C                  | 25.1 °C            | 0.109 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 32 °C                      | 31.746 °C                  | 31.9 °C            | 0.154 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 20 %HR                     | 18.935 %HR                 | 20.6 %HR           | 1.665 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 48.866 %HR                 | 50.4 %HR           | 1.534 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 68.182 %HR                 | 70.2 %HR           | 2.018 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |