

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.055 °C	15.2 °C	0.145 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.991 °C	25.2 °C	0.209 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.746 °C	31.9 °C	0.154 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	18.935 %HR	20.3 %HR	1.365 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	48.866 %HR	50.4 %HR	1.534 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	68.182 %HR	69.7 %HR	1.518 %HR	± 5.0 %HR	± 1.9E+00 %HR