

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.440 °C        | -25.1 °C  | 0.340 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| -15 °C            | -15.623 °C        | -15.2 °C  | 0.423 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.241 °C         | 0.2 °C    | 0.441 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |