

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.440 °C	-25.0 °C	0.440 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.241 °C	0.2 °C	0.441 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.851 °C	7.9 °C	0.049 °C	± 0.5 °C	± 4.0E-01 °C