

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.895 °C	2.6 °C	0.705 °C	± 0.5 °C	± 4.0E-01 °C
6 °C	5.826 °C	6.5 °C	0.674 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.851 °C	8.4 °C	0.549 °C	± 0.5 °C	± 4.0E-01 °C