

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.895 °C	1.9 °C	0.005 °C	± 0.5 °C	± 4.0E-01 °C
7 °C	6.979 °C	6.8 °C	-0.179 °C	± 0.5 °C	± 4.0E-01 °C
12 °C	11.932 °C	11.8 °C	-0.132 °C	± 0.5 °C	± 4.0E-01 °C