

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.440 °C	-25.2 °C	0.240 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-15.623 °C	-15.3 °C	0.323 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.241 °C	0.2 °C	0.441 °C	± 0.5 °C	± 4.0E-01 °C