

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
49.4 RPM	50 RPM	0.6 RPM	± 3 RPM	±	9.4E-02 RPM
49.6 RPM	50 RPM	0.4 RPM	± 3 RPM	±	9.4E-02 RPM
49.3 RPM	50 RPM	0.7 RPM	± 3 RPM	±	9.4E-02 RPM