

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
60.234 s	60.23 s	-0.004 s	-0.0066%	± 0.1%	±	5.0E-02 s
60.235 s	60.23 s	-0.005 s	-0.0083%	± 0.1%	±	5.0E-02 s
60.232 s	60.23 s	-0.002 s	-0.0033%	± 0.1%	±	5.0E-02 s