

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.042 °C	30.3 °C	0.258 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.787 %HR	31.8 %HR	0.013 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.000 %HR	52.1 %HR	1.100 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.967 %HR	73.1 %HR	-1.867 %HR	± 3.0 %HR	± 1.9E+00 %HR