

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.042 °C	30.5 °C	0.458 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.787 %HR	34.6 %HR	2.813 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.000 %HR	53.7 %HR	2.700 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.967 %HR	72.4 %HR	-2.567 %HR	± 3.0 %HR	± 1.9E+00 %HR