

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	14.931 °C	14.7 °C	-0.231 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.153 °C	29.8 °C	-0.353 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.787 %HR	33.1 %HR	1.313 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.867 %HR	69.4 %HR	-1.467 %HR	± 3.0 %HR	± 1.9E+00 %HR