

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	29.7 °C	-0.362 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	27.3 %HR	-2.963 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.611 %HR	48.9 %HR	-1.711 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	69.4 %HR	-0.945 %HR	± 5.0 %HR	± 1.9E+00 %HR