

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	29.6 °C	-0.462 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	25.2 %HR	-5.063 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.611 %HR	46.1 %HR	-4.511 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	66.8 %HR	-3.545 %HR	± 3.0 %HR	± 1.9E+00 %HR