

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
25 °C	24.848 °C	25.0 °C	0.152 °C	± 0.5 °C	± 4.0E-01 °C
# 25 °C	24.848 °C	25.0 °C	0.152 °C	± 0.5 °C	± 7.4E-02 °C
# 50 °C	49.917 °C	50.1 °C	0.183 °C	± 0.5 °C	± 7.4E-02 °C
# 70 °C	69.972 °C	70.2 °C	0.228 °C	± 0.5 °C	± 7.4E-02 °C
25 %HR	23.413 %HR	25.0 %HR	1.587 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	48.975 %HR	50.5 %HR	1.525 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	69.425 %HR	71.0 %HR	1.575 %HR	± 3 %HR	± 1.9E+00 %HR