

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
258.7 RPM	260 RPM	1.3 RPM	± 3 RPM	±	5.8E-02 RPM
963.4 RPM	960 RPM	-3.4 RPM	± 3 RPM	±	9.0E-02 RPM
1118.3 RPM	1120 RPM	1.7 RPM	± 3 RPM	±	9.0E-02 RPM