

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
149.12 RPM	150 RPM	0.88 RPM	± 3 RPM	±	1.1E-01 RPM
175.80 RPM	175 RPM	-0.80 RPM	± 3 RPM	±	5.8E-02 RPM
289.44 RPM	290 RPM	0.56 RPM	± 3 RPM	±	2.0E-01 RPM