

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-30 °C	-30.404 °C	-30.1 °C	0.304 °C	± 0.5 °C	± 4.0E-01 °C
-25 °C	-25.290 °C	-25.0 °C	0.290 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-14.962 °C	-15.0 °C	-0.038 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.048 °C	-10.0 °C	0.048 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.985 °C	2.0 °C	0.015 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.002 °C	8.0 °C	-0.002 °C	± 0.5 °C	± 4.0E-01 °C