

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-30 °C	-30.404 °C	-30.6 °C	-0.196 °C	± 0.5 °C	± 4.0E-01 °C
-25 °C	-25.290 °C	-25.4 °C	-0.110 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-14.962 °C	-15.2 °C	-0.238 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.048 °C	-10.2 °C	-0.152 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.985 °C	1.8 °C	-0.185 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.002 °C	8.1 °C	0.098 °C	± 0.5 °C	± 4.0E-01 °C