

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-30 °C	-30.404 °C	-29.8 °C	0.604 °C	± 0.5 °C	± 4.0E-01 °C
-25 °C	-25.290 °C	-24.8 °C	0.490 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-14.962 °C	-14.8 °C	0.162 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.048 °C	-9.7 °C	0.348 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.985 °C	2.4 °C	0.415 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.002 °C	8.3 °C	0.298 °C	± 0.5 °C	± 4.0E-01 °C