

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.222 °C	15.1 °C	-0.122 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.119 °C	25.0 °C	-0.119 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.012 °C	29.8 °C	-0.212 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	60.223 %HR	58.3 %HR	-1.923 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.131 %HR	63.5 %HR	-1.631 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68.6 %HR	-1.643 %HR	± 3.0 %HR	± 1.9E+00 %HR