

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
24.997 ml	25 ml	0.003 ml	± 2.0 ml	±	7.8E-01 µl