

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.1	0.09943	0.570	± 0.6	±	4.3E-01 mm						
0 a 0.5	0.49935	0.650	± 0.8	±	4.3E-01 mm						
0 a 1	0.99928	0.720	± 1	±	4.3E-01 mm						
0 a 2	1.99940	0.600	± 1.4	±	4.8E-01 mm						
0 a 3	2.99928	0.720	± 1.8	±	4.8E-01 mm						
0 a 4	3.99928	0.720	± 2.2	±	4.8E-01 mm						
0 a 5	4.99930	0.700	± 2.6	±	4.8E-01 mm						
0 a 6	5.99937	0.630	± 3	±	3.6E-01 mm						
0 a 7	6.99939	0.610	± 3.4	±	3.6E-01 mm						
0 a 8	7.99937	0.630	± 3.8	±	3.6E-01 mm						