

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.471 °C	28.6 °C	-0.871 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	33.750 %HR	32 %HR	-1.750 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	52.441 %HR	49 %HR	-3.441 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.284 %HR	69 %HR	-1.284 %HR	± 5.0 %HR	± 1.9E+00 %HR