

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.471 °C	28.8 °C	-0.671 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	33.750 %HR	37 %HR	3.250 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	52.441 %HR	56 %HR	3.559 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.284 %HR	75 %HR	4.716 %HR	± 5.0 %HR	± 1.9E+00 %HR