

Tipo de Medida <i>(Measurement Type)</i>	Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>
Profundidad / (Depth)	4.156 in	4.156 in	0 in	± 0.0019 in	± 3.6E-02 in
Profundidad / (Depth)	4.218 in	4.218 in	0 in	± 0.0019 in	± 3.6E-02 in
Profundidad / (Depth)	4.281 in	4.281 in	0 in	± 0.0019 in	± 3.6E-02 in
Profundidad / (Depth)	4.343 in	4.343 in	0 in	± 0.0019 in	± 3.6E-02 in
Profundidad / (Depth)	4.406 in	4.406 in	0 in	± 0.0019 in	± 3.6E-02 in
Profundidad / (Depth)	4.468 in	4.468 in	0 in	± 0.0019 in	± 3.6E-02 in