

Certificado No. (Certificate #):63076Equipo (Instrument):Medidor de Espesores

Fecha de Calibración (Calibration Date):2025-11-14Marca (Brand):VVV-Group

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
100.0 μm	101.0000 μm	1.0000 μm		±	2.5E+01 μm
197.0 μm	194.0000 μm	-3.0000 μm		±	2.5E+01 μm
295.0 μm	292.0000 μm	-3.0000 μm		±	2.5E+01 μm
400.0 μm	401.7500 μm	1.7500 μm		±	2.5E+01 μm
488.0 μm	494.7500 μm	6.7500 μm		±	2.5E+01 μm
595.0 μm	589.2500 μm	-5.7500 μm		±	2.7E+01 μm
900.0 μm	908.7500 μm	8.7500 μm		±	2.6E+01 μm
978.0 μm	994.0000 μm	16.0000 μm		±	2.5E+01 μm
1,485.0 μm	1,500.5000 μm	15.5000 μm		±	3.0E+01 μm