

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-18 °C	-18.802 °C	-18.5 °C	0.302 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.001 °C	4.0 °C	-0.001 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	54.282 %HR	54 %HR	0.018 %HR	± 3.0 %HR	± 1.9E+00 %HR