

Certificado No. (Certificate #):63090Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-11-14Marca (Brand):No Indicada

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.0000 mm	1.0000 mm	0.0000 µm	± 4 µm	±	5.8E-01 µm
2.0000 mm	2.0000 mm	0.0000 µm	± 4 µm	±	5.8E-01 µm
7.0000 mm	7.0010 mm	1.0000 µm	± 4 µm	±	5.8E-01 µm
8.0000 mm	8.0010 mm	1.0000 µm	± 4 µm	±	5.8E-01 µm
10.0000 mm	10.0010 mm	1.0000 µm	± 4 µm	±	5.8E-01 µm
12.0000 mm	12.0030 mm	3.0000 µm	± 4 µm	±	5.8E-01 µm
15.0000 mm	15.0030 mm	3.0000 µm	± 4 µm	±	5.8E-01 µm
18.0000 mm	18.0040 mm	4.0000 µm	± 4 µm	±	5.8E-01 µm
20.0000 mm	20.0030 mm	3.0000 µm	± 4 µm	±	5.8E-01 µm
22.0000 mm	22.0040 mm	4.0000 µm	± 4 µm	±	5.8E-01 µm
25.0000 mm	25.0040 mm	4.0000 µm	± 4 µm	±	5.8E-01 µm