

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.70 μm	52.74 μm	0.04 μm	± 1.6 μm	±	5.8E-02 μm
155.00 μm	154.8 μm	-0.20 μm	± 4.7 μm	±	6.2E-02 μm
257.00 μm	257.4 μm	0.40 μm	± 7.7 μm	±	7.0E-02 μm
512.00 μm	512.6 μm	0.60 μm	± 15.4 μm	±	9.7E-02 μm