

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.101 °C	0.3 °C	0.401 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.051 °C	15.4 °C	0.349 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.082 °C	30.1 °C	0.018 °C	± 0.5 °C	± 4.0E-01 °C
25 %HR	23.760 %HR	22.6 %HR	-1.160 %HR	± 5.0 %HR	± 1.9E+00 %HR
45 %HR	44.793 %HR	47.5 %HR	2.707 %HR	± 5.0 %HR	± 1.9E+00 %HR
65 %HR	65.110 %HR	69.7 %HR	4.590 %HR	± 5.0 %HR	± 1.9E+00 %HR