

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.361 °C	0.1 °C	0.461 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.025 °C	20.3 °C	0.275 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	34.795 °C	35.2 °C	0.405 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.660 %HR	29.7 %HR	0.040 %HR	± 5.0 %HR	± 1.9E+00 %HR
55 %HR	54.589 %HR	51.3 %HR	-3.289 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	79.957 %HR	75.8 %HR	-4.157 %HR	± 5.0 %HR	± 1.9E+00 %HR