

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.361 °C	-0.1 °C	0.261 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.025 °C	20.1 °C	0.075 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	34.795 °C	34.9 °C	0.105 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.660 %HR	31.4 %HR	1.740 %HR	± 5.0 %HR	± 1.9E+00 %HR
55 %HR	54.589 %HR	54.4 %HR	-0.189 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	79.957 %HR	80.9 %HR	0.943 %HR	± 5.0 %HR	± 1.9E+00 %HR