

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.074 ml	50 ml	-0.074 ml	± 0.5 ml	±	1.5E+00 µl
75.05 ml	75 ml	-0.052 ml	± 0.5 ml	±	2.3E+00 µl
125.08 ml	125 ml	-0.075 ml	± 1.0 ml	±	3.8E+00 µl