

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.039 ml	100 ml	-0.039 ml	± 0.5 ml	±	1.4E-01 µl
500.06 ml	500 ml	-0.056 ml	± 2.5 ml	±	1.4E-01 µl
999.97 ml	1000 ml	0.029 ml	± 5.0 ml	±	1.5E-01 µl