

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
99.989 ml	100 ml	0.011 ml	± 0.5 ml	±	1.6E-01 µl
499.96 ml	500 ml	0.043 ml	± 2.5 ml	±	1.4E-01 µl
999.99 ml	1000 ml	0.006 ml	± 5.0 ml	±	1.5E-01 µl