

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|------------|
| 1.0020 ml                | 1 ml             | -0.002 ml            | ± 0.05 ml                   | ±                                  | 7.8E-02 µl |
| 5.0018 ml                | 5 ml             | -0.002 ml            | ± 0.13 ml                   | ±                                  | 3.0E-01 µl |
| 10.0029 ml               | 10 ml            | -0.003 ml            | ± 0.12 ml                   | ±                                  | 5.9E-01 µl |