

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.0005 ml	1 ml	-0.001 ml	± 0.05 ml	±	7.8E-02 µl
5.0027 ml	5 ml	-0.003 ml	± 0.13 ml	±	3.0E-01 µl
10.0138 ml	10 ml	-0.014 ml	± 0.12 ml	±	5.9E-01 µl